



# EU Declaration of Conformity

In accordance with EN ISO 17050-1:2010



***The product manufacturer and location of Technical Construction File,***

Setra Systems, Inc  
159 Swanson Road, Boxborough, MA 01719, USA  
Tel: +1-978-263-1400 Fax: +1-978-264-0292

***and the authorized representative established within the Community,***

Hengstler GmbH  
Uhlandstr. 49, 78554 Aldingen, Germany  
Tel: +49-7424-890, Fax +49-7424-89500

***herewith declare that the products as originally delivered:***

*Product Name and Model:* Model 3100/3200 Pressure Transducer  
*Model number:* 310x, 320x (x=alphanumerical numbers for different configurations)  
*Options:* All

***comply with the essential requirements of the following applicable European Directives and carry the CE marking accordingly:***

EMC Directive (2014/30/EU)  
RoHS Directive (EU) 2015/863

***The following standards were used to demonstrate compliance:***

<b><i>EMC Standard</i></b>	<b><i>Limit</i></b>
CISPR 11:2009+A1:2010	Group 1, Class B Emission Limits
EN61326-1:2013 & EN61326-2-3: 2013	Industrial Immunity Requirements
EN 61000-4-2:2009	4 kV contact, 8 kV air
EN 61000-4-3:2006+A1:2008+A2:2010	10 V/m, 80 MHz - 1 GHz 3 V/m, 1.4 - 2 GHz 1 V/m, 2 - 2.7GHz 1kHz, 80% AM
EN 61000-4-4:2004+A1:2010	2 kV DC power lines, 1 kV signal/control lines
EN 61000-4-5:2006	2 kV Line-Ground, 1 kV Line-to-Line
EN 61000-4-6:2009	3V <sub>rms</sub> , 0.15-80MHz, 1 kHz, 80% AM

***DoC Number: CE-3100\_3200-E***

I hereby declare that the product(s) named above have been designed to comply with the relevant sections of the above referenced standards. The product(s) comply with all applicable Essential Requirements of the Directives. Subject to correct installation, maintenance and use conforming to its (their) intended purpose, to the applicable regulations and standards, to our operation and maintenance manual, the product complies with the provision of the above mentioned European Directives.

March 31, 2021

Date

**Eric Beishline**  
V.P. Engineering